



Ref. Certif. No.

JPTUV-144149

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Power Supply Unit
Name and address of the applicant	SELEC CONTROLS PVT LTD EL 27/1,ELECTRONIC ZONE TTC INDUSTRIAL AREA, MIDC MAHAPE NAVI MUMBAI, Maharashtra 400709, India
Name and address of the manufacturer	SELEC CONTROLS PVT LTD EL 27/1,ELECTRONIC ZONE TTC INDUSTRIAL AREA, MIDC MAHAPE NAVI MUMBAI, Maharashtra 400709, India
Name and address of the factory	Selec Controls Pvt Ltd EL30, Electronic Zone TTC Industrial Area MIDC Mahape Navi Mumbai Maharashtra 400710, India
Ratings and principal characteristics	AC INPUT: 100-240V~, 2.0-0.8A, 47-63Hz DC OUTPUT: +12Vdc - +48Vdc , 12.50A-3.13A 12V/0.5A (FAN), 150 WATT MAX. Class I (for OPS2x4-150-XXX-VCYY-1-CU), Class II (for OPS2x4-150-XXX-VCYY-2-CU)
Trademark (if any)	SELEC
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	refer to page 2
Additional information (if necessary may also be reported on page 2)	The risk management proces was excluded. For model differences, refer to the test report.
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1 See Test Report for National Differences
As shown in the Test Report Ref. No. which forms part of this Certificate	IN228A82 001

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland Japan Ltd.
Global Technology Assessment Center
4-25-2 Kita-Yamata, Tsuzuki-ku
Yokohama 224-0021, Japan
Phone + 81 45 914-3888
Fax + 81 45 914-3354
Mail: info@jpn.tuv.com
Web : www.tuv.com

Date: 2023-02-02

Signature:

Dipl.-Ing. (FH) B. Scheirer

Additional information :

Model:

OPS2x4-150-XXX-VCYY-Z-CU

where: "XX" denotes rated output voltage.

Third "X" denotes fractional value of output voltage (could be blank or any number, 1 to 9).

"V" denotes rated input voltage range (blank for 100-240 VAC).

"C" denotes connector type (any alphabet).

"YY" denotes minor output voltage variation (blank or any alphanumeric).

"Z" denotes class of product (could be 1 for Class I or 2 for Class II).



Date: 2023-02-02

Signature:

Dipl.-Ing. (FH) B. Scheirer